

CERTIFICATION

Helium – leak - check	SNIFFER - TEST
Date:	18 – 11 - 2008
Location:	high - vacuum area
Object:	Dichtungseinsatz C
Based on:	DIN EN 1779 Anhang B3
Charging pressure:	1000 mbar
First Test:	< 1·10⁻⁷ mbar·l/s
Test after 17 hours:	1,5·10⁻⁴ mbar·l/s

Terms of environmental:	Temperature :	18 °C	Humidity :	29 %
--------------------------------	---------------	--------------	------------	-------------

Calibration-standard:	Internal testleak			
	Reference N.:	-----		
	Date of calibration:	-----		

Helium-leakdetector:	Adixen			
	Modell :	ASM , 142 G		
	Serien Nr.:	30174		

Infraserv GmbH
Gleiwitzerstraße 8
85398 Eching

Geschäftsführer:
Dipl. Ing.
Christian Maenecke

Tel: 089 319 01 03
Fax: 089 31 9010 60

E-Mail:
info@infraservgmbh.com
www.infraservgmbh.com

Postfach Adresse:
Postfach: 11 44
85378 Eching

Bankverbindung:
Sparkasse Freising
BLZ: 700 510 03
Kto: 146 782 62

Dresdner Bank
BLZ: 700 800 00
Kto: 07841 851 00

Handelsregister München:
HRB 101844

Ust.-Id.-Nr.:
DE 156015582

Passed:
Hermann Scharl
Technical Director



Operator:
V.Hoffmann
Technician


